



EC type-examination certificate

Number **T10468** revision 0

Project number SO12200099

Page 1 of 1

Issued by NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

In accordance with The Metrologiewet (Stb. 2006, 137) as Dutch implementation of Directive 2004/22/EC on measuring instruments (MID).

Manufacturer Rocar-Tech Garage Equipment B.V.
Granaatstraat 55
7554 TN Hengelo
The Netherlands

In respect of A model of an **Exhaust Gas Analyzer**.
Type : 810

Characteristics Accuracy class 0
Electromagnetic environment class E1
Temperature range +5 °C / +40 °C

In the description number T10468 revision 0 further characteristics are described.

Valid until 3 May 2022

Description and documentation The instrument is described in the description number T10468 revision 0 and documented in the documentation folder T10468-1, appertaining to this EC-type examination certificate.

The Notified Body no. 0122
NMI Certin, 3 May 2012

C. Oosterman
Head Certification Board

NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see www.nmi.nl).

Reproduction of the complete document only is permitted.



1 General information about the exhaust gas analyzer

All properties of the exhaust gas analyzer, whether mentioned or not, may not be in conflict with the legislation.

1.1 Essential parts

| Description | Drawing number | Remark |
|-----------------------------------|----------------|---------------------------------|
| Block diagram of 810 gas Analyzer | 10468/0-01 | |
| Layout of control Circuit Board | 10468/0-02 | Drawing Parts list (1 page) |
| Layout of Microbench II | 10468/0-03 | Drawing Parts list (4 pages) |

| Oxygen sensor | | |
|---------------|-----------|--------|
| Manufacturer | Type | Remark |
| Tecnomotor | 1.61.0205 | - |

EMC protection measures:

- Power supply filter, Omega, type POLYSNAP BZ01110;
- Ferrite core around the temperature cable, 1 turn;
- Ferrite core around the RPM-counter cable, 1 turn.

1.2 Essential characteristics

Measuring ranges:

| Gas component | Display range | | Resolution | |
|-----------------|---------------|---------|------------|---------|
| CO | 0 – 10,0 | % vol | 0,01 | % vol |
| CO ₂ | 0 – 20,0 | % vol | 0,1 | % vol |
| HC | 0 – 10 000 | ppm vol | 1 | ppm vol |
| O ₂ | 0 – 21 | % vol | 0,01 | % vol |

Number **T10468** revision 0
Project number SO12200099
Page 2 of 4

The primary indication is on the 810. The firmware is according to guide WELMEC 7.2:

- Software type P;
- Risk Class C;
- No extensions are applicable.

The firmware is built according drawing:

- "Block diagram of the firmware", drawing number 10468/0-04.

| Firmware specification | | | |
|-----------------------------------|----------------|----------|-----------------|
| Firmware item | Version number | Checksum | Indication |
| Gas bench firmware | 5.129 | - | during start-up |
| Gas analyser firmware (English) | 2.2 | 1852 | during start-up |
| Gas analyser firmware (French) | 2.2 | 23280 | during start-up |
| Gas analyser firmware (Greek) | 2.2 | 1879 | during start-up |
| Gas analyser firmware (Italian) | 2.2 | 115 | during start-up |
| Gas analyser firmware (Polish) | 2.2 | 11900 | during start-up |
| Gas analyser firmware (Portugese) | 2.2 | 312 | during start-up |
| Gas analyser firmware (Romanian) | 2.2 | 1135 | during start-up |
| Gas analyser firmware (Spanish) | 2.2 | 671 | during start-up |
| Gas analyser firmware (Svedish) | 2.2 | 6041 | during start-up |
| Gas analyser firmware (Turkish) | 2.2 | 65470 | during start-up |

- Calibration period, 12 months.
- Functions:
 - Lambda calculation;
 - PEF in display;
 - (Semi-) automatic zero setting;
 - Drift compensation;
 - Low flow signaling.

- Protections:
 - Warm-up check;
 - Checking facilities for correct functioning;
 - Detection of HC residue;
 - Leak-test;
 - Control of O₂ channel;
 - Detection of adjustments that are necessary;
 - Signaling for ending of the calibration period;
 - Hardware sealing for the gas calibration.
- Parameters:

| | | |
|---------------------|---|----------------|
| - Warm-up time | : | < 10 min |
| - Temperature range | : | +5 °C / +40 °C |
| - Minimal flow | : | 1,5 l/min |
- Power supply voltage 230 V AC, 50 Hz.

1.3 Essential shapes

The exhaust gas analyzer is built according to the drawings:

- "Front view of the 810-830 Gas Analyzer", drawing number 10468/0-05;
- "Rear view of the 810-830 Gas Analyzer", drawing number 10468/0-06.

Markings:

- The markings have to fulfil the requirements stated in the legislation;
- The data plate is fixed to the exhaust gas analyzer and secured against removal by sealing or will be destroyed when removed;
- Near the display the inscriptions belonging to the function as exhaust gas analyzer must be present.

To secure components that may not be dismantled or adjusted by the user, the exhaust gas analyzer has to be secured in a suitable manner on the locations indicated in the drawing:

- "Security seal applied to gas tester", drawing number 10468/0-07.

1.4 Conditional parts

- User manual exhaust gas analyzer "Gas Analyzer 810-830 Operating manual, Revision 0.2", 52 pages;
- Pump, KNF, type PL3041A –NFT30;
- Valve, ASCO, type SCG356A001V;
- Gas anti-condensate filtering system, StampoTecnica, type 042.10.00039;
- Carbon Filters, FTI srl, type DIAN01;
- Power supply, Meanwell, type S150-13.5;
- Gas Probe, Tecnomotor.

The exhaust gas analyzer may be equipped with one or more of the following protective interfaces that have not to be secured:

- PS/2;
- USB;
- RS232.



Description

Number **T10468** revision 0
Project number SO12200099
Page 4 of 4

1.5 Non-essential parts

- Optional NO_x sensor;
- Revolutions counter;
- Oil temperature meter.

The exhaust gas analyzer may be connected to non-essential devices, for example but not limited to external printers, second display's, etc. provided that:

- They do not present primary data;
- They do not lead to an instrument having other essential characteristics than those fixed by this type-examination document.

2 Approval conditions, seals and verification marks

See chapter 1.2, essential characteristics and 1.3, essential shapes.