



Nederlands Meetinstituut

EC-type examination certificate

Number **T10102** revision 2
Project number 808846
Page 1 of 1

Issued by NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

Notified Body number 0122

In accordance with The Metrologiewet (Stb. 2006, 137) as Dutch implementation of Directive 2004/22/EC on measuring instruments (MID).

Manufacturer Robert Bosch GmbH
Robert-Bosch-Straße 10
73207 Plochingen
Germany

In respect of A model of an **Exhaust Gas Analyzer**.
Type : BEA 050 & BEA 460

Characteristics Accuracy class 0
Mechanical environment class M1
Electromagnetic environment class E2
Temperature range +5 °C / +40 °C

In the description number T10102 revision 2 further characteristics are described.

Valid until 30 January 2018

Description and documentation The instrument is described in the description number T10102 revision 2 and documented in the documentation folder T10102-1, appertaining to this EC-type examination certificate.

Remarks This revision replaces the earlier version, except for its documentation folder.

Dordrecht, 16 October 2008
NMI Certin B.V.

Ing. C. Oosterman
Manager Product Certification

NMI Certin BV
Hugo de Grootplein 1
3314 EG Dordrecht
PO Box 394
3300 AJ Dordrecht, NL
T +31 78 6332332
F +31 78 6332309
certin@nmi.nl
www.nmi.nl

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see "Regulation objection and appeal against decisions of NMI")

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

Reproduction of the complete document is permitted.



1 General information about the exhaust gas analyzer

All properties of the exhaust gas analyzer, whether mentioned or not, may not be in conflict with the legislation.

1.1 Essential parts

Description	Drawing number	Rev.	Remarks
Block diagram	Blockdiagramm BEA 050 Blockdiagramm BEA 460	- -	1 page 1 page
Pneumatic diagram	1 687 023 419	4	1 page
Main board	1 688 400 223 Stueckliste 1 688 400 223	1 -	layout, 1 page parts list, 3 pages
Board layout Andros 6510 gas bench	Model 6500/6510 Circuit Board Connector Locations Andros 6500 Bill of Materials Model 6510 Service and Repair Schedule	- - -	layout, 1 page parts list, 2 pages assembly, 1 page

Oxygen sensor		
Manufacturer	Type	Remarks
Teledyne	Bosch R17-A	part number 1 687 224 732
Envitec	Bosch A7-11.5	part number 1 687 224 843

1.2 Essential characteristics

Operation, indication and control:

- BEA 050:
 - Operation, indication and control via personal computer with CE marking. Communication between gas analyzer and personal computer via RS232 connection or optional via wireless BlueTooth connection;
- BEA460:
 - Operation, indication and control via personal computer with CE marking. Communication between gas analyzer and personal computer via wireless BlueTooth connection or usage of a RS232 cable.

Number **T10102** revision 2
Project number 808846
Page 2 of 4

Measuring ranges:

Gas component	Display range	Resolution
CO	0 – 10 % vol	0.001 % vol
CO ₂	0 – 18 % vol	0.01 % vol
O ₂	0 – 22 % vol	0.01 % vol
HC	0 - 9999 ppm vol	1 ppm vol

Software:

Software specification			
Software item	Version number	Checksum	Indication
Gas bench firmware	000-B6	\$F54B	MID-Screen
Gas analyzer firmware BEA050	2.01 or 2.02	\$30E5 \$3AA3	MID-Screen
Gas analyzer firmware BEA460	2.01 or 2.02	\$3BA3 \$3871	MID-Screen
PC program 1: EsaResEp.dll	1.0 or 1.10	\$3DA0 \$F04F	MID-Screen
PC program 2: EsaIO_Amo_Ep.dll	1.0 or 1.10	\$C37C \$C59D	MID-Screen
PC program 3: ScsCrc32.dll	1.0	\$CCF6	MID-Screen
PC program 4: WartAMM.exe	1.45 or 1.46	\$8872 \$856D	MID-Screen

The software versions and identification numbers will be displayed on the PC, in the ESA MID screen. The only legally relevant software is the MID part of the software (see module table above).

Re-calibration period:

- The calibration period is 12 months.

Functions:

- Lambda calculation;
- PEF in display;
- Automatic zero setting;
- Automatic calibration device;
- Drift compensation;
- Low flow signaling.

Number **T10102** revision 2

Project number 808846

Page 3 of 4

Protections:

- Warm-up check;
- Control of automatic devices;
- Detection of HC residue;
- Leak-test;
- Control of O₂ channel;
- Pump test;
- Filter test;
- Detection of adjustments that are necessary;
- Signaling for ending of the calibration period;
- Hardware sealing for firmware update;
- Password protection for gas calibration.

Parameters:

- Warm-up time : max. 5 min
- Temperature range : +5 °C / +40 °C
- Minimal flow : 140 l/h

Power supply voltage:

- The power supply voltage is 230 V AC, 50 Hz (optional: 110V AC, 60Hz).

1.3 Essential shapes

The exhaust gas analyzer is built according to the drawings:

- Front / rear view BEA050, drawing number 1 687 023 418, revision 1 of 7 May 2007;
- Front / rear view BEA460, drawing number 0 684 105 462 of 25 April 2007;
- Front / rear view AMM module, see "pneumatic diagram" drawing number 1 687 023 419, revision 4 of 31 May 2007.

Markings:

- The markings have to fulfill the requirements stated in the legislation;
- The data plate is fixed to the exhaust gas analyzer and secured against removal by sealing or will be destroyed when removed.

To secure components that may not be dismantled or adjusted by the user, the exhaust gas analyzer has to be secured in a suitable manner on the locations indicated in the drawings:

- Sealing BEA050, drawing number "Sealing BEA050";
- Sealing BEA460, drawing number "Sealing BEA460".

The securing component has to bear either:

- A mark of the manufacturer laid down in a by a Notified Body approved quality system, or;
- A mark of a Notified Body.

1.4 Conditional parts

- User manual:
 - Exhaust gas analyzer BEA050, dated 30 January 2008, 13 pages;
 - Exhaust gas analyzer BEA460, dated 25 January 2008, 15 pages;
 - ESA software, dated 28 November 2007, 85 pages;
- Pump, manufacturer, Rietschle Thomas Puchheim GmbH, Type 5002 VD / 5010 F;
- Filters, manufacturer, Mahle, type KL13;
- Gas probe, see drawing number 1 680 790 049, revision 4 of 24 March 2004 or drawing number 1 680 790 052, revision 3 of 24 March 2004;
- Motor measurement board (BEA 460 only) see drawing number 1 688 410 201.

The exhaust gas analyzer may be equipped with one or more of the following protective interfaces that have not to be secured:

- RS232 (1x);
- Centronics;
- PS/2 keyboard (not used with BEA050 and BEA460).

1.5 Non-essential parts

- Optional NO_x sensor.
- Integrated revolution counter based on:
 - Optical sensor;
 - Induction clamp;
 - UDP sensor;
 - Adaptor for dynamo ripple on board voltage;
- Oil temperature meter;
- OBD-Connector (BEA 460);
- RTM-Connector.

The exhaust gas analyzer may be connected to non-essential devices, for example but not limited to external printers, second display's, etc. provided that:

- They do not present primary data;
- They do not lead to an instrument having other essential characteristics than those fixed by this type-examination document.

2 Approval conditions, seals and verification marks

See chapter 1.2, essential characteristics and 1.3, essential shapes.